

"Your Total Patient Lift Solution"

EZ Way Accessory Inspection Checklist


The EZ Way slings and harnesses are made specifically for EZ Way lifts and stands. For the safety of the patient and caregiver, only EZ Way accessories should be used with EZ Way lifts and stands. Before operating the unit, ensure the accessory is not ripped, frayed or showing signs of wear.

Check binding and loops for any fraying, wear, nicks, or tears. Replace if fraying or wear is found.

> **EXAMPLES OF ACCESSORIES** THAT NEED REPLACEMENT







Check sling or harness body fabric for any rips, holes, fraying, or weak spots. Hold sling up to a light. The sling must be replaced if any light shows through.



Check all stitching of strap connections, (Box-X stitching) for loose stitches. Check all binding stitching for loose stitches. Pull on all straps/loops in opposite direction and note if stitching become loose or comes apart. Remove from service if stitching tears.





Check handles for loose stitches and tears. Remove from service if any found.





Check all straps where they attach to the sling hanger bars for wear. Remove from service if any fraying is found.



EZ Way offers a 6-month warranty on slings and harnesses and recommends replacement after one year or if the sling or if the sling/harness shows any sign of damage or wear.

PO Box 89, 710 E. Main. Clarinda, IA 51632 | 1-800-627-8940 | sales@ezlifts.com

WARNING: Cancer and Reproductive Harm - www.P65Warnings.ca.gov.