



EZ Way, Inc.

"Your Total Patient Lift Solution"

EZ 4000/4200

Atlas Bariatric Shower Trolleys

Operator's Instructions



EZ Way, Inc.
PO Box 89
Clarinda, IA 51632
1-800-627-8940
www.ezlifts.com

Form 2-200
Rev. 07/31/18

⚠WARNING: Cancer and Reproductive Harm - www.P65Warnings.ca.gov.

The EZ 4000/4200 Atlas Bariatric Multi Tasking Shower Trolley is designed for lots of power and loads of space. The trolley is used for showering, transport and as a changing or treatment trolley.

The EZ 4000/4200 Shower Trolley makes showering and transferring easier for attendants and patients.

Safety Notes

The patient must be positioned comfortably on the stretcher mattress with side rails SECURELY locked. Never leave the patient unattended.

Check that the side rails are locked in place and free from damage.

Do not let the patient sit on the foldable short head end rail on the stretcher.

Caster brakes must be securely locked on the trolley before transferring a patient and once in position over a drain. Check that braking attachments and casters are clean and free from damage.

Operate the trolley carefully when transporting patients through narrow corridors or over irregular uneven surfaces.

Ensure that the trolley can be moved, raised or lowered without risking any pinching injuries to patient.

Trolley must be cleaned after each use. Wipe underneath the stretcher to prevent water from dripping on the floor. Never clean mattress and pillow in gas sterilization unit. Mattress and pillow will explode.

Stretcher and mattress must be kept in place, well maintained and all screws must be tightened.

Trolley must be inspected visually for signs of wear. Trolley should not be used if defects are noted. If defects are noticed, prompt service is required.

Only use an EZ Way-approved battery charger for the trolley. Do not use battery charger in wet or damp areas as this is a potential shock hazard. It is recommended that the trolley batteries are charged daily or whenever trolley is not in use. Batteries cannot be overcharged.

Always study this manual carefully before operating the EZ 4000/4200 Shower Trolleys. Operating instructions and recommended maintenance procedures are explained in the following pages.

Only qualified staff who understand all operating and safety procedures may operate this equipment. The EZ Shower Trolleys are designed for years of trouble free operation when properly used and maintained.

If you have any questions regarding the operation or maintenance of your EZ 4000/4200 Shower Trolley, contact us at 1-800-627-8940. Additional copies of these instructions can be downloaded from EZ Way's website, www.ezlifts.com.

WARNING:

Do not use the shower trolley in any way that varies from the safety precautions and operating procedures detailed in this manual. The equipment must not be operated by anyone until the user has read and understood the instructions in this manual.

Features and Controls

Position	Description
1.	Flexible drainage hose
2.	Drainage
3.	Mattress
4.	Foldable side rail
5.	Foldable head rail for extra length (only for EZ 4000 Atlas)
6.	Foot control brakes and straight steering
7.	Bumper wheel
8.	Column for height adjustments
9.	Straight steering caster
10.	Wedge pillow 15° (option)
11.	Wedge pillow 30° (option)
12.	Hand control
13.	Up
14.	Down
15.	Battery charger input
16.	Tilting





Figure 1 - Drain Hose, Plug and Hand Control



Figure 2 - 4000 Shower Trolley

Audible Tone Sounds

Short Tone

This tone informs the user that the stretcher is level when tilting and stops. Proceed tilting by pressing once more on the hand control when desired.

Long and Short Tone

When this tone sounds, the trolley needs to be charged.

Long and Two Short Tones

If/when this tone sounds, the stretcher actuators must be calibrated to level.

Calibrating Actuators

Use the hand control to calibrate by pressing the DOWN symbol until the stretcher stops in its lowest position. Press the DOWN symbol once more to make sure that both actuators are in their lowest positions.

Two short audible tones sound when both actuators are in their lowest position. Press the DOWN symbol once more and the stretcher is now calibrated and level.

Calibration must also be performed after each power failure, for example after changing the batteries or the fuse.

Hand Controls for Up/Down and Tilting




Figure 3 - Hand Control

The stretcher is easily and quietly raised by two powerful actuators. Load limit is up to 1000 lb. for the EZ 4000 Atlas and 600 lb. for the EZ 4200 Atlas Junior.

The hand control operates the power when activated and stops immediately when released. The hand control also operates the tilting in both

directions and raises / lowers the stretcher quietly.

Raise the shower trolley stretcher by activating the "UP" arrow symbol. Lower the shower trolley stretcher by activating the "DOWN" arrow symbol.

When needed, the stretcher can be tilted to a maximum of 4° in both directions by activating the tilt head- foot symbols on the hand control for improved drainage and to a Trendelenburg position.

NOTE: Only qualified staff who understands the procedures may operate the trolley with a patient in a "Trendelenburg" position.

Shower Trolley Rails

The long side rails swing down for patient transfer and lock in up position for patient security.

Lower the Side Rail (Horizontal) (Figure 4)

Lift the side rail with both hands and lower it to a horizontal position where it stops.



Figure 4 - Side Rail Lowered Partially

Lower the Side Rail (Completely) (Figure 5)

Lift the side rail with both hands from the horizontal position and lower it all the way.



Figure 5 - Side Rail Lowered Completely

Raise the Side Rail

Raise the side rail with both hands by simply lifting it back into its upright position and it will automatically lock in place.

The stretcher width can be narrowed as the side rails can be folded inwards when transporting the trolley through narrow doorways.

Head End Rail (Only EZ 4000 Atlas)

The head end rail lowers down to a horizontal position for extra length and more access when transferring.

Lower the End Rail (Figure 6)

Lift rail with both hands and lower.



Figure 6 - End Rail Lowered

Raise the Head End Rail

Raise the head end rail with both hands by simply lifting it to its upright position where it automatically locks in place.

Corner Bumper Wheels

The stretcher has rubber bumper wheels in all four corners to protect walls and doorways.

Central Locking Caster Brakes & Straight Steering




Fig. 7 - Central Locking Control Bar Positions

Caster Brakes

Park the shower trolley and brake all four casters by pushing the single foot central locking control bar, under the trolley head end, all the way down until it clicks in position (Figure 8).



Figure 8 - Brake Position

Release the brakes on all four casters by lifting the single foot central locking control bar to its neutral position (Figure 9), or push down the short ends on the foot control bar.



Figure 9 - Neutral Position

Straight Steering Caster

One of the foot end casters can be locked in a straight steering position. Lock the caster by lifting the foot central locking control bar until it clicks to position or push down short ends on the foot control bar (Figure 10).



Figure 10 - Locking Position

To release the straight steering caster, push the foot central locking control bar down to its neutral position (Figure 9), or lift the short ends on the foot control bar.

Drain Hose and Plug

Drain Hose

The flexible drain hose connects to the drain outlet in the mattress, which prevents splashing. To reduce water dripping on the floor after use, fold the drain hose nozzle upwards.

Drainage Plug

The drainage plug is fastened with a nylon string on the stretcher foot end rail.



Figure 11 - Drain Hose And Plug

Patient Handling Procedure

Patient Showering

1. When showering, the patient must remain on the shower trolley stretcher with the side rails securely locked.
2. Always physically ensure that the water temperature is as required before and during showering.
3. Park the shower trolley foot end over a floor drain, sink or WC. Brake all four casters by pushing the single foot central locking control bar all the way down, until it clicks in position (D).
4. When desired, attach the drainage hose and drain plug.
5. Tilt the stretcher max 4 degrees, with the hand control for a comfortable position and good drainage. Remove the drain plug, if desired.
6. Shower the patient. Body fluids and feces must be removed immediately to prevent mattress stains.
7. After the patient is showered keep him/her warm wrapped in a bath towel.
8. Thoroughly drain the stretcher and attach the drain plug to prevent water from dripping on floor. Fold the drain hose nozzle upwards to prevent water dripping on the floor.
9. When desired, level the stretcher with the hand control. An audible short tone sounds when the stretcher is back in level.

NOTE: When the patient requires the "Trendelenburg" position, only qualified staff who understands the procedures may operate the trolley.

NOTE: Never leave patient unattended.

Cleaning the Shower Trolley

The powder coating protects the shower trolley from corrosion. It is sealed against moisture and can be washed and cleaned thoroughly and easily.

The stretcher has a removable PVC coated mattress and pillow (option). They must always be cleaned after each use.

The mattress and pillow cloth are heat sealed to prevent moisture penetration.

Remove and clean the support shelves on studs on the stretcher frame.

Rinse the stretcher, support shelves, mattress and pillow thoroughly with the patient shower handle after cleaning and let dry.

NOTE: Wipe underneath shower trolley to prevent water from dripping on floor.

WARNING: Never clean the mattress and / or pillow in a gas sterilization unit. Mattress and pillow will explode.

Battery Power and Charging

- Charge the trolley, in dry rooms only, daily or when a long and a short audible tone sounds. Normal charging time is approximately 8 hours when flat. The trolley can never be overcharged.
- Use only the 24 volt DC battery charger model delivered by EZ Way.
- Connect the battery charger into the shower trolley hand control input. Make sure that the plug is fully in place.
- Plug the battery charger into the mains.
- When the battery charger is properly connected, an indicator on the battery charger will be lit. If not, recheck the connections.
- The indicator tells the charging cycles. A yellow light emits during charging and green indicates that the trolley is ready for use.
- When the charging is completed, disconnect the charger from the mains before you disconnect the charger from the hand control.
- A flat fuse (15 A) is connected between the batteries in the battery box underneath the stretcher.

Batteries

If changing the lead batteries, deposit in accordance with local regulations.



Figure 12 - Battery charger




Figure 13 - Hand Control

Routine Maintenance

Patient Showering

The recommended frequency of these maintenance routine procedures will depend on how often the equipment is used.

It is good practice to establish these procedures and then adjust the frequency as your conditions indicate.

Daily Routine Maintenance

- Clean the trolley, mattress and pillow after each use with mixed, nonabrasive type of detergent, according to your detergent supplier instructions.
NOTE: Do not use solvent based cleaning liquids, which can damage or stain the PVC coating on the mattress and pillow.
- Use a sponge or brush, and remove any soap or lime deposits, and feces promptly to avoid spots/stains. Rinse thoroughly with fresh water.
- It is recommended that the trolley batteries are charged daily or whenever trolley is not in use. Batteries can not be overcharged.

Weekly Maintenance

- Visually inspect the shower trolley interior to make sure there are no nicks or cracks which may scratch the patient.

Monthly Maintenance

- Check that all screws and fastenings are tight.
- Check that the siderails and casters are working properly.
- Inspect shower and cleaning equipment.

WARNING: Never clean mattress pad and pillow in gas sterilization unit. Pad and pillow will explode.

Batteries

If changing the lead batteries, deposit in accordance with local regulations.

WARRANTY, CE MARK & TECHNICAL SPECIFICATIONS:

EZ Way 4000/4200 Atlas Bariatric Shower Trolleys

Warranty: EZ Way warrants that its products are free from defects in materials and workmanship for a period of one year from the date of delivery. EZ Way mattresses and pillow of fabric covered PVC over polyester foam shall be warranted for a period of six months from the date of delivery. This warranty is in lieu of any other warranty or representation, expressed or implied. Warranties of MERCHANTABILITY or FITNESS FOR ANY PARTICULAR USE OR PURPOSE are expressly excluded. EZ Way will not be responsible for any expenses, inconvenience or incidental or consequential damages occasioned by the use, loss of use, or misuse of the products, or by the failure of the products to conform to the warranty set forth above. This warranty becomes void if the product shows evidence of mishandling, tampering, installation or use contrary to the applicable instruction materials, shipping damage, or repair. Routine cleaning, adjustment, and normal cosmetic and mechanical wear are not covered by this warranty.



CE mark: Certificate of registration at the LÄKEMEDELSVERKET Medical products agency in Sweden
Dnr.452:2008/83584.

Free sales certificate at the LÄKEMEDELSVERKET Medical products agency in Sweden.

TECHNICAL SPECIFICATIONS

	EZ 4000 Atlas	EZ 4200 Atlas Junior
Length:	87"	88"
Width:	36"	29"
Height:	21" - 33"	21" - 33"
Working Load Limit:	1,000 lbs.	530 lbs.
Chassis:	Powder-coated and zinc-plated steel	Powder-coated and zinc-plated steel
Casters:	Brakes on all four castors and straight steering	Brakes on all four castors and straight steering
Control:	Hand control for height adjustment and tilting	Hand control for height adjustment and tilting
Mattress, Pillow:	Flame-retardant material	Flame-retardant material
Battery:	Rechargeable, 24 V	Rechargeable, 24 V
Twin Actuators:	Battery Powered	Battery Powered
Intermittent Power:	10% max 2 min. / 18 min.	10% max 2 min. / 18 min.